

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年06月


	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/6/1	3.6	4.4	4.2	4.6	4.7	5.1	-20.1	-19.7	-79.9	-79.5	18.7	19.1
2019/6/2	3.6	4.4	4.2	4.6	4.7	5.1	-20.1	-19.8	-79.9	-79.6	18.6	18.7
2019/6/3	2.8	4.7	4.1	4.7	4.7	5.3	-20.7	-18.9	-81.2	-79.4	19	19.8
2019/6/4	3.4	4.5	4.1	4.7	4.7	5.3	-20.9	-20.4	-81.4	-78.9	18.8	19.9
2019/6/5	3.4	4.5	4.1	5.2	4.6	5.2	-20.9	-20	-80.4	-78.7	18.8	19.8
2019/6/6	2.8	4.9	4.2	4.8	4.6	5.3	-20.7	-20	-80.4	-78.9	18.8	20.1
2019/6/7	3.3	4	4.2	4.7	4.8	5.3	-20.7	-20.2	-80.4	-78.6	18.7	19.1
2019/6/8	3.3	4.1	4.2	4.7	4.8	5.2	-20.7	-20.2	-80.4	-78.8	18.6	18.7
2019/6/9	3.3	4.1	4.2	4.7	4.8	5.2	-20.7	-20.2	-80.4	-78.9	18.6	18.6
2019/6/10	2.9	4	4.2	4.8	4.5	5.3	-20.8	-20.2	-80.4	-78.6	18.4	19.6
2019/6/11	3.2	4.2	4.2	4.9	4.5	5.3	-20.9	-20.3	-80.4	-78.6	18.7	19.8
2019/6/12	3.4	5.7	4.2	4.7	4.6	5.3	-20.8	-20.2	-80.3	-78.9	18.8	19.8
2019/6/13	3.7	5.4	4.2	4.9	4.7	5.3	-20.7	-19.9	-80.4	-77.9	18.8	19.8
2019/6/14	3.7	5.5	4.3	4.8	4.7	5.3	-20.7	-20.1	-79.6	-76.4	18.8	19.7
2019/6/15	3.7	4.3	4.3	4.8	4.8	5.3	-20.7	-20.3	-79.4	-77.6	19.1	19.2
2019/6/16	3.7	4.3	4.3	4.8	4.8	5.2	-20.7	-20.3	-79.3	-77.8	19.1	19.2
2019/6/17	3.7	5.6	4.2	4.8	4.7	5.3	-20.7	-19.9	-79.3	-77.7	19.1	20.1
2019/6/18	3.7	5.2	4.2	4.7	4.7	5.3	-20.4	-19.7	-79.4	-76.8	19.2	20.4
2019/6/19	3.9	5.4	4.2	4.7	4.7	5.3	-20.7	-19.9	-79.5	-74.7	19.4	20.6
2019/6/20	3.9	5.7	4.2	4.8	4.4	5.3	-20.7	-19.2	-79.6	-75.8	19	20.1
2019/6/21	4.1	5.5	4.2	4.7	4.3	5.1	-20.2	-19.3	-79.6	-77.7	18.9	19.9
2019/6/22	4	4.4	4.2	4.6	4.5	5.1	-20.3	-19.8	-79.6	-77.8	18.7	19.1
2019/6/23	4	4.4	4.2	4.6	4.4	5.1	-20.3	-19.9	-79.6	-77.5	18.6	18.7
2019/6/24	4.1	5.4	4.2	4.7	4.3	5	-20.6	-19.6	-79.6	-77.7	18.4	19.6
2019/6/25	4.1	5.4	4.2	4.7	4.4	5	-20.6	-19.9	-79.6	-77.8	18.6	19.7
2019/6/26	4.1	5.6	4	4.8	4.4	5.1	-20.4	-19.9	-79.6	-77.7	18.8	19.7
2019/6/27	3.8	5.8	4	4.9	4.5	5.2	-20.4	-19.6	-79.6	-77.6	18.7	19.8
2019/6/28	3.8	5.4	4.1	4.9	4.5	5.1	-21.2	-19.6	-79.6	-77.6	18.8	19.7
2019/6/29	3.8	4.3	4.1	4.6	4.7	5.2	-21.2	-20.7	-79.5	-77.8	18.8	19.4
2019/6/30	3.8	4.3	4.1	4.6	4.7	5.1	-21.2	-20.7	-79.6	-77.8	18.7	19.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內


 藥學部 邱保祥 108/6/10
 藥學部 溫雅如 108/6/10
 師二如 溫雅如 108/6/10