

# 臺北榮民總醫院藥學部

## 臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 108年06月

	六號冰箱		室溫	
序號	1436-0086		1118-0011	
備用	0436-049		0742-0169	
範圍	2~8°C		15~25°C	
日期	Min	Max	Min	Max
2019/6/1	4	4.4	19.2	19.4
2019/6/2	4	4.4	18.8	19.2
2019/6/3	3.9	4.5	19.2	19.7
2019/6/4	3.7	4.4	19.3	19.7
2019/6/5	3.8	4.4	19.3	19.6
2019/6/6	3.9	4.4	19.3	20.2
2019/6/7	3.9	4.4	19.4	19.8
2019/6/8	3.9	4.4	19.2	19.4
2019/6/9	3.9	4.4	19.1	19.3
2019/6/10	3.9	4.4	19	19.7
2019/6/11	4	4.4	19.4	19.9
2019/6/12	3.8	4.5	19.4	20.1
2019/6/13	3.9	4.4	19.5	20
2019/6/14	3.9	4.7	19.5	20.1
2019/6/15	4.1	4.5	19.7	20.1
2019/6/16	4.1	4.5	19.3	19.9
2019/6/17	3.9	4.5	19.1	19.5
2019/6/18	3.6	4.7	19.2	20.1
2019/6/19	3.9	4.4	19.2	20.2
2019/6/20	3.9	4.5	19.4	20.2
2019/6/21	4	4.4	19.4	19.9
2019/6/22	4	4.4	19.4	19.7
2019/6/23	4	4.4	19.3	19.5
2019/6/24	3.6	4.5	19.4	19.8
2019/6/25	3.6	4.2	19.6	20.1
2019/6/26	3.5	4.2	19.8	20.2
2019/6/27	3.5	4.1	19.7	20.6
2019/6/28	3.6	4.2	19.6	20.3
2019/6/29	3.6	4.1	19.6	19.8
2019/6/30	3.6	4.2	19.4	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 108  
師(三)級藥師 0101 藥學部 溫雅如 108  
師(三)級藥師 0101

溫度均在正常範圍內