

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 108年09月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1213-0010		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1405-0193		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2019/9/1	4.3	4.7	4.3	4.8	4.5	5.1	-19.1	-18.9	-81.4	-79.7	18.2	18.6
2019/9/2	4	4.7	4.3	4.8	4	4.9	-20.2	-19.1	-81.3	-79.7	18.8	19.5
2019/9/3	4.1	4.8	4.3	5.1	4	4.9	-20.2	-19.7	-81.3	-79.7	18.7	19.6
2019/9/4	4.2	4.8	4.4	5.1	4.1	5	-19.9	-18.9	-81.4	-79.7	18.8	19.8
2019/9/5	4.2	5.1	4.4	5.1	4.1	5.1	-19.4	-19	-81.4	-79.2	18.8	19.8
2019/9/6	4.1	4.9	4.4	5.1	4	4.9	-19.2	-18.8	-81.4	-79.7	18.7	19.9
2019/9/7	4.2	4.7	4.5	5.1	4.5	5	-19.2	-19.1	-81.5	-79.4	18.4	18.9
2019/9/8	4.2	4.7	4.5	5	4.5	4.9	-19.3	-19	-81.5	-79.8	18.1	18.4
2019/9/9	4.2	4.9	4.4	5.1	4	4.9	-19.6	-19.1	-81.4	-79.8	18.2	19.6
2019/9/10	4.3	4.8	4.3	5.3	4.1	4.9	-19.6	-19	-81.4	-79.2	18.6	19.7
2019/9/11	4.3	4.9	4.3	5.1	4.1	4.9	-20.2	-19.1	-81.4	-79.6	18.8	19.8
2019/9/12	4.3	5.1	4.3	5.1	4.2	5.1	-20.3	-19.6	-81.4	-79.7	18.8	19.9
2019/9/13	4.4	4.9	4.5	5	4.8	5.2	-20.2	-19.9	-81.3	-79.3	18.8	19.1
2019/9/14	4.5	4.8	4.4	5	4.7	5.1	-20.2	-19.9	-81.4	-79.3	18.6	18.8
2019/9/15	4.4	4.8	4.4	5	4.7	5.1	-20.2	-19.9	-81.5	-79.5	18.2	18.6
2019/9/16	4.3	4.9	4.3	5.1	4.4	5.2	-20.7	-19.9	-81.5	-79.7	17.9	19.7
2019/9/17	4.3	4.9	4.3	5.1	4.4	5.2	-20.7	-19.8	-81.4	-79.4	19.1	20.4
2019/9/18	4.4	4.9	4.3	5.1	4.4	5.2	-21.1	-19.4	-81.2	-79.7	19.3	20.4
2019/9/19	4.4	5.3	4.4	5.4	4.1	5.1	-20.9	-20.1	-81.2	-79.5	19.4	20.4
2019/9/20	4.3	5.1	4.4	5.2	4.1	5.1	-20.9	-19.6	-81.2	-79.7	19.5	20.6
2019/9/21	4.5	5.1	4.7	5.2	4.5	4.9	-20.9	-19.8	-81.4	-79.2	19.6	19.9
2019/9/22	4.4	5	4.7	5.2	4.5	4.9	-21.4	-20.9	-81.2	-79.7	19.4	19.6
2019/9/23	4.2	4.9	4.5	5.2	4.2	5.1	-23.1	-20.5	-81.2	-79.7	19.3	20.3
2019/9/24	4.2	4.8	4.6	5.2	4.4	5.1	-22.1	-19.3	-81.2	-79.5	19.4	21
2019/9/25	4.2	4.8	4.6	5.3	4.3	5.1	-20.4	-19.4	-81.3	-79.7	19.5	20.4
2019/9/26	4.2	4.7	4.5	5.3	4.3	5.1	-21.1	-19.7	-81.3	-79.7	19.4	20.5
2019/9/27	4.2	5.4	4.5	5.3	4.2	5.1	-20.6	-19.9	-81.4	-79.3	19.4	20.3
2019/9/28	4.2	4.7	4.7	5.3	4.8	5.2	-20.4	-20.1	-81.4	-79.8	19.3	19.6
2019/9/29	4.2	4.7	4.7	5.3	4.8	5.1	-20.4	-20.2	-81.4	-79.7	19.2	19.3
2019/9/30	4.2	4.7	4.7	5.3	4.8	5.1	-20.4	-20.1	-81.3	-79.5	19.2	19.4

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 108  
師(三)級藥師 邱保祥 1001 藥學部 溫雅如 108  
師(二)級藥師 溫雅如 1001