

臺北榮民總醫院藥學部

臨床試驗藥局藥品冷藏/冷凍/室溫溫度管制表 106年 3月

	一號冰箱		二號冰箱		三號冰箱		四號冰箱		負20度冰箱		負80度冰箱		室溫	
序號	0520-362		0520-293		1504-0305		1611-0052		1213-0010		1215-0006		1213-0007	
備用	0580-0835		0520-292		1504-0339		1611-0062		1405-0193		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2017/3/1	4.9	5.6	3.8	5.2	4.5	5.6	4.9	5.4	-20.3	-18.8	-80.8	-79.8	19.5	20.6
2017/3/2	4.9	5.7	3.9	5.2	4.3	5.4	4.9	6.3	-20.4	-19.7	-80.8	-79.7	19.2	20.5
2017/3/3	4.9	5.8	3.9	5.2	4.2	5.4	5.3	5.8	-20.2	-19.6	-80.9	-79.7	19.2	20.7
2017/3/4	5.1	5.8	4.3	5.2	5.1	5.4	5.5	5.9	-19.9	-19.8	-80.9	-79.8	19.4	19.7
2017/3/5	5.1	5.9	4.3	5.2	4.9	5.4	5.3	5.9	-20.2	-19.7	-80.8	-79.8	17.5	20.1
2017/3/6	4.9	5.8	4.1	5.2	4.3	5.4	5.3	5.9	-20.2	-19.6	-80.9	-79.7	18.1	21.1
2017/3/7	4.9	5.7	3.7	5.2	4.4	5.5	5.2	5.8	-20.1	-19.8	-80.9	-79.7	19.7	21.4
2017/3/8	4.8	5.7	3.8	5.1	4.3	5.4	4.2	5.7	-20.2	-19.8	-80.8	-79.7	19.8	21.7
2017/3/9	4.8	5.7	3.8	5	4.2	5.4	4.2	4.9	-20.3	-19.9	-80.8	-79.8	19.7	21.3
2017/3/10	4.7	5.7	3.9	5.1	4.4	5.7	4.2	4.9	-20.4	-19.8	-80.8	-79.8	19.2	20.6
2017/3/11	4.8	5.7	4.1	5.2	5.3	5.7	4.2	4.8	-20.4	-20.2	-80.9	-79.8	18.7	19.4
2017/3/12	4.9	5.7	4.1	5.2	5.3	5.7	4.2	4.8	-20.6	-20.2	-80.9	-79.8	18.4	19.4
2017/3/13	4.5	5.7	4.1	5.2	4.5	5.7	4.2	4.8	-20.6	-19.9	-80.9	-79.8	18.3	19.6
2017/3/14	4.7	5.7	4.1	5.2	4.4	5.7	4.1	4.8	-20.2	-19.8	-80.9	-79.8	18.4	19.6
2017/3/15	4.6	5.6	4.2	5.2	4.4	5.6	4.1	4.9	-20.2	-19.9	-80.9	-79.8	18.3	19.7
2017/3/16	4.7	5.6	4.2	5.2	4.4	5.7	4.2	4.8	-20.2	-18.8	-80.9	-79.8	17.8	20
2017/3/17	4.6	5.7	4.2	5.3	4.4	5.8	4.1	4.9	-20.2	-19.7	-80.9	-79.8	18.4	20.1
2017/3/18	5.1	5.7	4.2	5.3	5.3	5.7	4.2	4.8	-20.3	-19.9	-80.9	-79.9	17.9	18.9
2017/3/19	5.1	5.7	4.2	5.3	5.3	5.6	4.2	4.8	-20.3	-20.1	-80.9	-79.9	18.3	18.4
2017/3/20	4.6	5.8	4.2	5.3	4.4	5.6	4.2	4.9	-20.7	-20.1	-80.9	-79.8	18.3	19.7
2017/3/21	4.6	5.7	4.1	5.4	4.5	5.7	4.2	5	-20.7	-20.4	-80.9	-79.8	18.6	20.4
2017/3/22	4.9	5.8	3.9	5.4	4.6	5.7	4.2	4.9	-20.7	-20.2	-80.9	-79.7	19.2	20.6
2017/3/23	4.8	5.8	3.9	5.2	4.4	5.7	4.2	4.9	-20.6	-20.2	-80.9	-79.9	18.9	20.3
2017/3/24	4.8	5.8	3.9	5.2	4.3	5.4	4.3	4.9	-20.4	-20	-80.9	-79.8	18.3	19.6
2017/3/25	5.1	5.9	4	5.1	5.1	5.5	4.3	4.9	-20.4	-20.1	-80.9	-79.8	18.8	19.6
2017/3/26	5.2	5.8	3.9	5.1	5.1	5.4	4.3	4.9	-20.4	-20.1	-80.9	-79.8	19	19.5
2017/3/27	4.8	5.9	3.8	5.2	4.3	5.4	4.2	4.9	-20.4	-19.9	-80.9	-79.8	18.8	20.6
2017/3/28	4.8	5.8	3.7	5.2	4.4	5.6	4.2	4.9	-20.4	-20.1	-80.9	-79.8	18.7	20.1
2017/3/29	4.8	5.8	3.7	5.1	4.5	5.7	4.3	4.9	-20.6	-20	-81.1	-79.8	19.2	20.8
2017/3/30	4.7	5.8	3.7	5.1	4.4	5.7	4.2	4.9	-20.6	-20	-80.9	-79.7	19.2	21.6
2017/3/31	4.8	5.8	3.5	5.1	4.4	5.7	4.1	4.8	-20.5	-19.9	-80.9	-79.8	19.2	20.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

藥學部 邱保祥
師(三)級藥師

藥學部 陳盈如
師(三)級藥師