

臺北榮民總醫院藥學部特殊製劑藥局  
臨床試驗藥品冷藏/室溫溫度管制表 106 年 3 月

|           | 一號冰箱                 |     | 室溫                   |      |
|-----------|----------------------|-----|----------------------|------|
| 序號        | Serial No. 1436-0086 |     | Serial No. 1118-0011 |      |
| 校正日期      | 2016/12/16           |     | 2016/12/16           |      |
| 範圍        | 2 - 8 °C             |     | 15 - 25 °C           |      |
| 日期        | Min                  | Max | Min                  | Max  |
| 2017/3/1  | 3.3                  | 4.1 | 19.3                 | 19.7 |
| 2017/3/2  | 3.4                  | 4.0 | 19.4                 | 19.8 |
| 2017/3/3  | 3.3                  | 4.2 | 19.3                 | 19.7 |
| 2017/3/4  | 3.6                  | 4.3 | 16.9                 | 17.9 |
| 2017/3/5  | 3.6                  | 4.7 | 17.7                 | 18.1 |
| 2017/3/6  | 4.1                  | 4.9 | 17.8                 | 18.2 |
| 2017/3/7  | 3.7                  | 4.5 | 17.3                 | 17.8 |
| 2017/3/8  | 4.1                  | 5.1 | 17.6                 | 17.9 |
| 2017/3/9  | 4.2                  | 5.8 | 17.3                 | 18.1 |
| 2017/3/10 | 4.3                  | 4.9 | 17.9                 | 18.8 |
| 2017/3/11 | 4.3                  | 4.7 | 18.2                 | 18.6 |
| 2017/3/12 | 4.3                  | 5.4 | 18.2                 | 18.6 |
| 2017/3/13 | 4.2                  | 4.5 | 18.4                 | 19.1 |
| 2017/3/14 | 4.2                  | 4.7 | 17.8                 | 18.6 |
| 2017/3/15 | 3.8                  | 4.7 | 17.6                 | 18.2 |
| 2017/3/16 | 3.9                  | 4.3 | 17.8                 | 18.6 |
| 2017/3/17 | 3.9                  | 4.5 | 18.4                 | 19.1 |
| 2017/3/18 | 4.0                  | 4.4 | 18.6                 | 18.8 |
| 2017/3/19 | 4.0                  | 4.4 | 18.6                 | 18.9 |
| 2017/3/20 | 4.0                  | 4.4 | 18.7                 | 19.2 |
| 2017/3/21 | 3.9                  | 4.9 | 18.2                 | 19.1 |
| 2017/3/22 | 3.9                  | 4.4 | 18.2                 | 19.3 |
| 2017/3/23 | 3.8                  | 4.8 | 18.6                 | 19.3 |
| 2017/3/24 | 3.8                  | 4.3 | 18.2                 | 19.1 |
| 2017/3/25 | 3.8                  | 4.3 | 18.8                 | 19.2 |
| 2017/3/26 | 3.8                  | 4.3 | 18.2                 | 18.9 |
| 2017/3/27 | 3.8                  | 4.3 | 17.7                 | 18.2 |
| 2017/3/28 | 3.6                  | 4.3 | 17.9                 | 18.6 |
| 2017/3/29 | 3.7                  | 4.2 | 18.3                 | 18.7 |
| 2017/3/30 | 3.7                  | 4.2 | 18.6                 | 20.0 |
| 2017/3/31 | 3.7                  | 4.3 | 19.2                 | 19.5 |

型號 Model No. EI-IN-D-32-L 溫度均在正常範圍內

負責藥師 藥學部 邵啟鈞 藥學部 詹徽明  
師(三)級藥師 師(三)級藥師