

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 04月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362												
0850-0835												
0520-292												
1504-0305												
1405-0193												
1215-0006												
1213-0010												
1330-0064												
1405-0194												
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/4/1	3.7	4.7	4	4.4	3.8	4.4	-20.9	-19.2	-80.7	-79.3	18.2	18.6
2024/4/2	3.7	4.7	3.9	4.4	3.8	4.4	-20.9	-19.9	-80.7	-79.3	18.1	18.8
2024/4/3	3.7	4.4	4	4.7	3.6	4.5	-20.2	-19.4	-80.7	-78.8	18.2	18.8
2024/4/4	3.7	4.3	4.3	4.6	3.7	4.2	-19.9	-19.7	-80.7	-79.3	18.2	18.4
2024/4/5	3.7	4.3	4.3	4.6	3.7	4.2	-19.9	-19.7	-80.8	-78.4	18.2	18.7
2024/4/6	3.7	4.2	4.3	4.7	3.6	4.2	-19.9	-19.6	-80.8	-79	18.7	18.8
2024/4/7	3.7	4.2	4.3	4.6	3.7	4.2	-19.9	-19.7	-80.7	-78.7	18.7	18.8
2024/4/8	3.7	4.4	4.3	4.9	3.6	4.2	-19.8	-19.4	-80.7	-78.9	18.7	19.3
2024/4/9	3.9	4.7	4.3	4.7	3.6	4.2	-19.8	-19.4	-80.8	-79.4	18.8	19.4
2024/4/10	3.6	5.1	4.1	4.7	3.6	4.3	-19.9	-19.5	-80.9	-77.7	18.9	19.6
2024/4/11	3.6	4.6	4.1	4.5	3.8	4.4	-19.9	-19.6	-80.8	-79.4	18.9	19.4
2024/4/12	3.6	4.5	4.1	4.6	3.8	4.3	-19.9	-19.6	-80.8	-76.2	19	19.6
2024/4/13	4	4.4	4.1	4.4	3.8	4.3	-19.8	-19.6	-80.7	-77.9	19	19.2
2024/4/14	4	4.4	4.1	4.4	3.8	4.3	-19.8	-19.6	-80.7	-77.8	18.8	19
2024/4/15	3.9	4.6	4.1	4.5	3.8	4.3	-20.9	-19.6	-80.7	-78.4	18.8	19.6
2024/4/16	4	4.7	4	4.5	3.5	4.3	-20.9	-20.1	-80.7	-78.7	19	19.6
2024/4/17	4	4.8	3.9	4.4	3.5	4.3	-20.4	-19.7	-80.7	-78.5	19	19.6
2024/4/18	3.9	4.7	3.9	4.6	3.1	4.3	-19.9	-19.6	-80.7	-79.3	19.1	19.6
2024/4/19	3.9	4.9	4	4.5	3.6	4.4	-19.9	-18.8	-80.7	-78.1	19.2	19.7
2024/4/20	4	4.5	4.1	4.4	3.7	4.2	-19.6	-19	-80.7	-79.4	19.2	19.3
2024/4/21	4	4.5	4.1	4.4	3.7	4.2	-19.7	-19.4	-80.6	-79.3	19.1	19.3
2024/4/22	4	4.6	4	4.4	3.7	4.2	-20.2	-19.3	-80.7	-77.9	19	19.5
2024/4/23	3.9	4.7	4	4.4	3.7	4.3	-20.7	-19.4	-80.7	-77.7	19	19.5
2024/4/24	3.7	5	4.1	4.6	3.7	4.3	-20.7	-20	-80.6	-78.1	19.2	19.5
2024/4/25	3.7	5.7	4.2	4.5	3.7	4.7	-20.7	-19.6	-80.7	-79.4	18.2	19.4
2024/4/26	3.9	4.6	4.2	4.7	3.6	4.3	-20.2	-19.6	-80.6	-78.3	17.9	18.9
2024/4/27	3.9	4.4	4.3	4.6	3.6	4.2	-19.7	-19.5	-80.7	-79.4	18.8	18.9
2024/4/28	3.9	4.4	4.3	4.6	3.6	4.2	-19.7	-19.6	-80.6	-79.3	18.8	18.8
2024/4/29	3.8	4.5	4.3	4.6	3.6	4.2	-19.8	-19	-80.6	-79.3	18.7	19.2
2024/4/30	3.8	4.6	4.2	4.7	3.5	4.2	-20.1	-19.2	-80.7	-78.2	18.9	19.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥
師(三)級藥師 113
0501 藥學部 邱子芹
師(三)級藥師 113
0801

總藥師: 藥學部 廖志峰
師(二)級藥師 0501
1576

業管主管: 藥學部 何沁沁
師(一)級藥師 0502
0830