

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113 年 04 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/4/1	5.5	6.1	3.7	4.6	4.7	5.1	3.9	4.7	4.4	4.9	19.3	19.7
2024/4/2	5.8	6.1	3.7	4.5	4.6	5.1	3.3	4.8	4.4	5.3	19.6	19.9
2024/4/3	5.7	6	3.7	4.6	4.7	5.3	3.3	4.8	4.9	5.4	19.6	19.9
2024/4/4	5.7	5.9	3.7	4.6	4.9	5.2	3.5	4.7	4.9	5.4	19.7	19.9
2024/4/5	5.7	5.9	3.8	4.6	4.9	5.2	3.5	4.8	4.9	5.4	19.5	19.8
2024/4/6	5.7	5.9	3.8	4.5	4.9	5.2	3.5	4.7	4.9	5.4	19.7	19.9
2024/4/7	5.7	5.9	3.7	4.6	4.7	5.2	3.5	4.7	4.9	5.4	19.8	19.8
2024/4/8	5.6	6.1	3.7	4.6	4.5	5.1	3.5	4.8	4.7	5.3	19.8	20.1
2024/4/9	5.6	5.9	3.7	4.6	4.2	5.2	3.2	4.8	4.7	5.4	20.1	20.2
2024/4/10	5.3	5.8	3.7	4.7	4.1	4.7	3.3	4.7	4.4	5.2	20.1	20.3
2024/4/11	5.3	5.9	3.9	4.7	4.1	4.6	3.3	4.7	4.4	4.9	20.2	20.4
2024/4/12	5.5	6.1	3.9	4.6	4.1	4.6	3.3	4.7	4.4	4.9	20.3	20.6
2024/4/13	5.7	5.9	3.9	4.6	4	4.6	3.3	4.7	4.4	4.9	20.4	20.6
2024/4/14	5.7	5.9	3.9	4.6	4.1	4.6	3.3	4.7	4.4	4.9	20.2	20.4
2024/4/15	5.7	6	3.9	4.7	4	4.8	3.3	4.8	4.4	5.1	20.1	20.5
2024/4/16	5.7	6	3.9	4.7	3.9	4.6	3.2	4.7	4.3	4.9	20.3	20.6
2024/4/17	5.7	5.9	3.8	4.5	4	4.5	3.2	4.7	4.3	4.8	20.4	20.6
2024/4/18	5.7	6	3.8	4.7	4.1	4.5	3.1	4.7	4.3	5.2	20.5	20.7
2024/4/19	5.7	6	3.7	6.1	4	5.3	3.3	4.7	4.4	5.1	20.5	20.7
2024/4/20	5.7	5.9	3.7	4.6	4.1	4.5	3.3	4.6	4.4	4.9	20.4	20.6
2024/4/21	5.7	5.9	3.7	4.6	4	4.5	3.3	4.7	4.4	4.9	20.2	20.4
2024/4/22	5.6	5.9	3.6	4.7	3.9	4.6	3.3	4.6	4.4	4.9	20.1	20.4
2024/4/23	5.6	5.8	3.6	4.8	3.9	5.4	3.5	4.7	4.4	5.3	20.2	20.5
2024/4/24	5.5	5.8	3.7	4.5	3.9	6.1	3.7	4.8	4.5	5.1	20.3	20.6
2024/4/25	5.4	5.7	3.6	4.6	3.9	4.6	3.2	4.7	4.5	5.1	19.9	20.5
2024/4/26	5.4	5.9	3.7	4.5	3.8	4.5	3.3	4.7	4.4	5.1	19.6	20.1
2024/4/27	5.3	5.6	3.7	4.5	3.9	4.4	3.3	4.6	4.4	4.9	20.1	20.1
2024/4/28	5.3	5.6	3.7	4.5	3.9	4.4	3.3	4.7	4.4	4.9	19.9	20.1
2024/4/29	5.3	6.1	3.7	4.7	3.9	4.4	3.3	4.7	4.4	4.9	19.8	20.1
2024/4/30	5.6	6	3.7	4.6	3.9	4.4	3.2	4.7	4.4	5.5	20.1	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 113
師(三)級藥師 0501 藥學部 邱子萍 113
師(三)級藥師 54

總藥師: 藥學部 廖志峰
師(二)級藥師 0570
1570

業管主管: 藥學部 何沁沁 0802
師(一)級藥師 0830