

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113 年 05 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/5/1	5.6	5.8	3.5	4.7	4	4.4	3.3	4.8	4.5	5.1	20.1	20.4
2024/5/2	5.6	6.6	3.5	4.5	3.8	4.6	3.4	4.9	4.5	5.1	20.3	20.7
2024/5/3	6.1	6.3	3.5	4.5	3.9	4.5	3.5	4.9	4.4	5	20.4	20.8
2024/5/4	6	6.3	3.6	4.7	3.9	4.4	3.6	5	4.4	5.1	19.8	20.7
2024/5/5	5.9	6.2	3.7	4.7	4	4.5	3.7	5.1	4.4	4.9	19	19.8
2024/5/6	6	6.3	3.6	4.7	3.9	4.5	3.5	4.9	4.4	4.9	19	19.8
2024/5/7	6.1	6.6	3.6	4.5	3.9	4.5	3.3	4.9	4.4	5.4	19.8	20.2
2024/5/8	6	6.3	3.6	4.6	3.9	4.5	3.3	4.9	4.4	4.9	19.9	20.3
2024/5/9	6	6.7	3.6	4.6	2.7	5.4	3.4	5.1	4.2	5	19.9	20.2
2024/5/10	6.4	6.6	3.6	5.2	2.7	4.2	3.7	4.9	4.2	4.8	19.9	20.2
2024/5/11	6.3	6.4	3.7	4.4	2.7	4.2	3.7	5	4.3	4.8	19.9	20.1
2024/5/12	6.3	6.4	3.7	4.4	2.8	4.3	3.7	5.1	4.3	4.8	19.7	19.9
2024/5/13	6.2	6.4	3.7	4.6	2.8	4.8	3.7	5.1	4.3	4.9	19.6	19.8
2024/5/14	6.1	6.5	3.7	4.6	4	5.6	3.1	5.1	4.5	4.9	19.5	20.1
2024/5/15	5.9	6.5	3.7	4.5	4.2	5.1	3.1	4.7	4.4	4.9	20.1	20.4
2024/5/16	6	6.2	3.7	4.6	4.1	4.9	3.3	4.7	4.4	5.4	20.2	20.5
2024/5/17	5.3	6.2	3.8	4.9	4.1	4.9	3.3	4.7	4.7	5.3	20.4	20.8
2024/5/18	5.3	5.6	4.3	4.9	4.2	4.8	3.6	4.8	4.7	5.2	20.6	20.7
2024/5/19	5.3	5.6	4.3	4.9	4.2	4.9	3.6	4.9	4.7	5.2	19.7	20.6
2024/5/20	5.3	6	4.3	4.9	3.6	4.8	3.6	5.1	4.7	5.2	19.9	20.3
2024/5/21	5.2	5.9	4.3	4.9	3.1	5.4	3.9	5.1	4.7	5.4	20.2	20.4
2024/5/22	5.2	5.6	4.4	5	2.9	4.3	4.1	5.1	2.4	6	20.2	20.6
2024/5/23	5.1	5.4	4.4	5.5	3	4.2	3.6	5	2.4	3.9	20.4	20.8
2024/5/24	5.1	5.5	3.9	5.1	3.1	5.2	3.6	5.1	2.6	6.1	20.3	20.8
2024/5/25	5.1	5.3	3.9	4.7	4.6	5.1	3.7	5.1	5.3	5.7	19.8	20.3
2024/5/26	5.1	5.3	3.9	4.7	4.6	5.1	3.7	5.1	5.3	5.7	20.2	20.3
2024/5/27	5.1	5.4	3.8	4.7	4.6	5.2	3.7	5.1	5.3	5.9	20.1	20.5
2024/5/28	5.1	5.3	3.8	4.8	4.6	5.1	3.9	5.4	5.6	6.1	20.1	20.4
2024/5/29	4.9	5.8	4.1	5.1	4.5	5.1	4.2	5.3	5.9	6.2	19.8	20.1
2024/5/30	5.1	5.5	4	4.8	4.4	4.9	4.2	5.2	5.5	6.2	19.9	20.2
2024/5/31	5.1	5.4	3.9	4.9	4.4	5.2	2.9	5.2	5.8	6.4	20.1	20.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 邱保祥 113
師(三)級藥師 0603

藥學部 邱子芹 113
師(三)級藥師 0604

總藥師:

藥學部 廖志峰
師(二)級藥師 0608
1720

業管主管:

藥學部 何沁沁 0605
科主任 0940