

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 06月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007							
0850-0835	0520-293	1504-0339	1213-0010	1330-0064	1405-0194							
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/6/1	4	4.7	2.7	3.8	3.4	4.1	-20.3	-20	-81.8	-79.9	18.3	18.6
2024/6/2	4	4.7	2.7	3.9	3.4	4.2	-20.3	-20	-81.8	-79.3	18.2	18.8
2024/6/3	3.2	4.5	2.9	4.1	3.7	4.7	-20.7	-17.4	-81.8	-79.8	18.8	19.2
2024/6/4	2.9	4.2	2.9	4	3.7	4.4	-20.7	-19.6	-81.8	-78.6	18.8	19.2
2024/6/5	2.9	4	3.1	4	3.7	4.3	-20.4	-18.8	-81.8	-79.9	18.6	19.9
2024/6/6	3.3	4.4	3.2	4	3.7	4.2	-19.6	-18.8	-81.8	-78.9	18.4	18.9
2024/6/7	3.4	4.5	3.2	3.9	3.6	4.4	-20.9	-19.6	-81.9	-79.7	18.2	18.8
2024/6/8	3.6	4.6	3.2	3.8	3.7	4.2	-20.9	-20.6	-81.9	-80.2	18.2	18.3
2024/6/9	3.6	4.6	3.2	3.8	3.7	4.2	-20.9	-20.6	-81.9	-79.4	18	18.2
2024/6/10	3.6	4.6	3.2	3.9	3.6	4.2	-20.9	-20.6	-81.9	-80.2	17.9	19.4
2024/6/11	3.6	4.6	3.2	4.1	3.7	4.2	-21.1	-20.6	-81.8	-79.8	17.9	18.7
2024/6/12	3.6	4.4	3.2	4	3.7	4.4	-21.2	-20.6	-81.8	-80.1	18.2	18.7
2024/6/13	3.3	4.3	3.2	4	3.8	4.2	-21.2	-20.9	-81.8	-78.8	17.5	18.9
2024/6/14	3.4	4.3	3.1	4.6	3.7	4.3	-21.4	-20.3	-81.8	-79.9	17.2	18.9
2024/6/15	3.5	4.2	3.1	3.7	3.7	4.2	-21.2	-20.8	-82	-80.2	18.9	19.1
2024/6/16	3.5	4.2	3.1	3.7	3.7	4.2	-21.2	-20.8	-81.9	-80.2	18.4	19.1
2024/6/17	3.6	4.3	3.1	3.9	3.8	4.4	-21.4	-20.8	-81.9	-79.8	17.8	18.4
2024/6/18	3.6	4.5	2.9	3.8	3.8	4.3	-21.4	-19.9	-81.8	-80.2	17.7	18.1
2024/6/19	3.7	4.4	2.9	3.8	3.9	4.3	-20.2	-19.2	-81.8	-79	17.7	18.4
2024/6/20	3.7	4.4	3.2	4.5	3.9	4.4	-19.6	-19	-81.8	-78.8	17.7	18.7
2024/6/21	3.7	4.3	3.7	4.3	3.9	4.3	-20.4	-19.4	-81.8	-79	17.7	18.2
2024/6/22	3.7	4.3	3.8	4.3	3.9	4.3	-20.4	-20.2	-81.8	-79.3	17.7	18.2
2024/6/23	3.7	4.3	3.8	4.3	3.9	4.3	-20.4	-20.2	-81.8	-79	17.7	18.1
2024/6/24	3.7	4.4	3.6	4.3	3.9	4.7	-20.4	-19.2	-81.8	-80.2	17.7	18.2
2024/6/25	3.7	4.4	3.8	4.3	4.2	4.7	-19.8	-18.7	-81.9	-80.2	17.7	18.1
2024/6/26	3.7	4.3	3.7	4.4	4.3	4.9	-19.5	-18.8	-81.9	-80.2	17.7	18.1
2024/6/27	2.6	4.4	3.7	4.2	4.3	4.7	-19.4	-18.6	-82	-80.2	17.7	18.1
2024/6/28	2.8	4.3	3.7	4.3	4.2	5.3	-19.6	-17.1	-81.9	-80.2	17.7	18.2
2024/6/29	3.9	4.2	3.7	4.3	4.4	4.7	-19.5	-19.1	-81.9	-79.7	17.4	18.1
2024/6/30	3.9	4.2	3.7	4.2	4.4	4.7	-19.4	-19.2	-81.9	-78.8	17.1	17.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 0701
藥學部 邱子芹 0702

總藥師: 藥學部 廖志峰 0703

業管主管: 藥學部 何沁沁 0704
科主任 1430

溫度均在正常範圍內