

臺北榮民總醫院藥學部

臨床試驗藥局細胞毒性藥品冷藏/室溫溫度管制表 113 年 06 月

	六號冰箱		七號冰箱		室溫	
序號	1436-0086		2123-0026		1118-0011	
備用	0436-049		2123-0027		0742-0169	
範圍	2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max
2024/6/1	4.4	4.9	4.4	5.3	21.9	22.3
2024/6/2	4.4	5	4.7	5.3	21.9	22.2
2024/6/3	4.3	5.7	4.3	5.5	19.8	22
2024/6/4	4.2	4.9	4.3	4.9	21.7	21.9
2024/6/5	4.2	4.9	4.4	4.9	21.7	22.1
2024/6/6	4.3	4.9	4.4	4.9	21.8	22.2
2024/6/7	4.2	5	4.2	4.9	21.8	22.3
2024/6/8	4.3	4.9	4.2	4.9	21.8	22.3
2024/6/9	4.3	4.9	4.2	4.8	21.7	22.2
2024/6/10	4.2	4.9	4.2	4.9	21.5	22.1
2024/6/11	4.3	4.9	4.2	4.9	21.7	22.2
2024/6/12	4	5.2	4	4.8	21.8	22.3
2024/6/13	4	4.8	4.2	4.9	21.9	22.4
2024/6/14	4	4.8	4.2	4.9	22.2	22.8
2024/6/15	3.9	4.7	4.4	4.9	22.3	22.9
2024/6/16	3.9	4.8	4.4	4.9	22.3	22.7
2024/6/17	3.9	4.8	4.4	5.3	22.2	22.8
2024/6/18	3.9	4.8	4.4	4.9	22.3	23
2024/6/19	4	4.8	4.5	5.2	22.6	23.2
2024/6/20	4.1	4.9	4.7	5.3	22.6	23.4
2024/6/21	4.1	4.8	4.7	5.3	22.6	23.2
2024/6/22	4.1	4.8	4.6	5.1	22.6	23.4
2024/6/23	4.1	4.8	4.6	5.1	22.7	23.3
2024/6/24	4.1	4.8	4.6	5.3	22.4	23.3
2024/6/25	4.1	4.8	4.7	5.3	22.3	22.9
2024/6/26	4.1	4.8	4.4	5.3	22.3	23.1
2024/6/27	4.1	5.1	4.7	5.5	22.6	23.2
2024/6/28	4.1	4.8	4.7	5.3	22.4	22.9
2024/6/29	4.1	4.8	4.7	5.3	22.2	22.6
2024/6/30	4.1	4.7	4.7	5.3	22	22.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 113 0701
藥學部 邱予芹 113 0702

總藥師: 藥學部 廖志峰 113 0702

業管主管: 藥學部 何沁沁 0704 1430