

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 07月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/7/1	4	4.5	3.5	4.3	4.1	5.5	-21.4	-20.2	-81.8	-80.1	17.3	18.4
2024/7/2	4	4.5	3.5	4.2	3.9	4.4	-21.4	-20.7	-81.8	-80.1	18.2	18.8
2024/7/3	3.9	4.9	3.5	4.3	3.9	4.4	-21.3	-20.8	-81.8	-80.1	18.1	18.9
2024/7/4	3.8	4.5	3.6	4.3	4	4.5	-21.2	-20.6	-81.8	-80.1	17.3	18.1
2024/7/5	3.8	4.4	3.5	4.3	4.1	4.5	-21.2	-20.7	-81.8	-79.3	17.2	17.9
2024/7/6	3.8	4.2	3.7	4.2	4.2	4.5	-21.2	-20.9	-81.9	-79.3	17.1	17.5
2024/7/7	3.8	4.2	3.7	4.2	4.2	4.5	-21.2	-20.9	-81.8	-78.6	16.9	17.4
2024/7/8	3.8	4.5	3.7	4.3	4.1	4.5	-21.4	-20.8	-81.8	-79.9	16.9	18.3
2024/7/9	3.6	4.3	3.7	4.3	3.9	4.6	-21.2	-20.9	-81.8	-80.1	17.7	18.3
2024/7/10	3.8	4.5	3.6	4.3	3.9	4.4	-21.2	-20.4	-81.8	-79.9	17.7	18.6
2024/7/11	3.7	5.5	3.6	4.2	3.9	4.3	-21.4	-20.6	-81.8	-80	18.2	18.9
2024/7/12	2.9	4.4	3.6	4.2	3.9	4.4	-21.4	-20.8	-81.9	-79.8	18.4	18.9
2024/7/13	2.9	3.9	3.6	4.2	3.9	4.3	-21.2	-20.9	-81.8	-80.2	18.4	18.6
2024/7/14	2.9	3.9	3.7	4.2	3.9	4.3	-21.2	-20.9	-81.9	-80.1	17.4	18.4
2024/7/15	3.1	4.3	3.7	4.4	4	4.5	-21.2	-20.4	-81.9	-80.2	17.2	18.2
2024/7/16	3.2	4.3	3.7	4.3	4.2	4.6	-20.9	-20.3	-81.8	-80.1	17.6	18.1
2024/7/17	3.3	4.3	3.7	4.4	4.2	4.7	-21.2	-20.4	-81.8	-80.1	17.7	18.2
2024/7/18	3.4	4.3	3.7	4.3	4.2	4.6	-21.2	-20.6	-81.8	-80.1	17.7	18.5
2024/7/19	3.3	4.3	3.7	4.3	4.2	4.9	-21.4	-20.9	-81.8	-79.9	17.7	18.3
2024/7/20	3.4	4.2	3.7	4.2	4.2	4.5	-21.4	-21.2	-81.8	-78.4	16.9	17.8
2024/7/21	3.4	4.2	3.7	4.3	4.2	4.5	-21.4	-21.1	-81.9	-78.7	16.9	17.3
2024/7/22	3.3	4.5	3.7	4.6	4.2	5.1	-21.4	-19.7	-81.8	-80.2	16.8	18.1
2024/7/23	3.5	4.4	3.7	4.3	4	4.4	-21.2	-20.7	-81.8	-80.2	17.5	18.2
2024/7/24	3.5	4.2	3.7	4.2	4	4.4	-21.2	-20.9	-81.8	-79.3	17.7	18
2024/7/25	3.5	4.3	3.7	4.2	4	4.4	-21.2	-20.9	-81.8	-79.3	17.7	17.9
2024/7/26	3.4	4.6	3.7	4.3	3.7	4.4	-21.2	-20.4	-81.8	-79.4	17.6	18.1
2024/7/27	3.3	4.3	3.9	4.3	3.8	4.2	-20.8	-20.4	-81.8	-80	17.5	17.7
2024/7/28	3.3	4.3	3.9	4.3	3.8	4.2	-20.8	-20.5	-81.8	-79.9	17.3	17.5
2024/7/29	3	4.4	3.8	4.3	3.8	4.3	-20.8	-20.5	-81.8	-80.1	17.3	19.1
2024/7/30	3.7	5.3	3.7	4.3	3.7	4.4	-20.7	-20.1	-81.8	-79.8	19.1	20.2
2024/7/31	4.1	4.8	3.7	4.3	3.6	4.4	-20.9	-20.2	-81.8	-79.3	19.5	19.9

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 13 0801
藥學部 邱子芹 13 0802

總藥師: 藥學部 廖志峰 14 20

業管主管: 藥學部 何沁沁 13 0806

溫度均在正常範圍內