

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113 年 07 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/7/1	3.9	4.2	3.4	4.6	3.9	4.7	4.5	5.2	4.6	5.3	19.3	20.1
2024/7/2	3.9	4.3	3.4	4.6	4	4.9	4.4	5.1	4.7	5.4	20.1	20.4
2024/7/3	3.9	4.3	3.3	4.7	4.1	4.8	4.4	5.1	4.6	5.5	20.3	20.5
2024/7/4	3.9	4.2	3.3	4.7	4	4.8	4.5	5	4.7	5.9	19.7	20.3
2024/7/5	3.7	4.2	3.4	4.6	4.2	5.3	4.5	5.1	4.7	5.5	19.5	19.8
2024/7/6	3.7	4.1	3.4	4.7	4.2	4.9	4.6	5	4.7	5.3	19.4	19.6
2024/7/7	3.7	4.1	3.4	4.7	4.2	4.9	4.6	5.1	4.7	5.3	19.2	19.4
2024/7/8	3.7	4.1	3.4	4.7	4	4.9	4.5	5.1	4.7	5.3	19.2	19.9
2024/7/9	3.7	4.1	3.6	4.7	3.9	4.8	4.5	5	4.7	5.3	19.7	20.1
2024/7/10	3.6	4.1	3.5	4.7	3.9	4.8	4.4	5.1	4.7	5.5	19.9	20.3
2024/7/11	3.6	4.2	3.4	4.6	3.9	4.8	4.5	5.1	4.8	5.4	20.2	20.5
2024/7/12	3.8	4.3	3.4	4.8	4	4.8	4.5	5.1	4.6	5.4	20.3	20.6
2024/7/13	3.8	4.1	3.7	4.7	4	4.8	4.5	5.1	4.5	5.2	20.3	20.5
2024/7/14	3.8	4.1	3.7	4.7	4	4.8	4.6	5.1	4.5	5.2	19.7	20.3
2024/7/15	3.8	4.2	3.7	4.8	3.9	4.8	4.6	5.1	4.5	5.2	19.4	19.8
2024/7/16	3.9	4.3	3.6	4.7	3.9	4.8	4.4	5.1	4.5	5.3	19.5	19.9
2024/7/17	3.9	4.2	3.6	4.8	3.9	4.9	4.4	4.9	4.7	5.3	19.7	19.9
2024/7/18	3.9	4.2	3.4	4.8	4.2	4.8	4.4	5.1	4.7	5.3	19.7	20.1
2024/7/19	3.9	4.2	3.5	4.7	4.2	4.8	4.4	5	4.6	5.4	19.7	20.1
2024/7/20	3.9	4.2	3.6	4.7	4.2	4.8	4.4	5	4.7	5.3	19	19.8
2024/7/21	3.9	4.2	3.6	4.7	4.2	4.8	4.4	5.1	4.7	5.2	18.8	19.1
2024/7/22	3.8	4.2	3.6	4.7	4.2	4.9	4.5	5.1	4.5	5.2	18.7	19.4
2024/7/23	3.8	4.2	3.6	4.7	4.4	5.2	4.5	5.1	4.5	5.2	19.4	19.7
2024/7/24	3.8	4.2	3.6	4.7	4.4	4.9	4.5	5.1	4.8	5.2	19.6	19.8
2024/7/25	3.8	4.2	3.5	4.7	4.3	4.9	4.6	5.1	4.7	5.2	19.6	19.8
2024/7/26	3.8	4.2	3.6	4.7	4.3	4.9	4.5	5.1	4.7	5.3	19.5	19.7
2024/7/27	3.9	4.2	3.7	4.7	4.3	4.9	4.6	5.1	4.7	5.3	19.3	19.6
2024/7/28	3.9	4.2	3.7	4.7	4.3	4.9	4.6	5.1	4.7	5.3	19	19.3
2024/7/29	3.9	4.2	3.7	4.7	4.3	5.1	4.6	5.1	4.7	5.3	18.9	19.9
2024/7/30	3.7	4.2	3.5	4.7	4.4	5.1	4.6	5.1	4.7	5.7	19.9	21.1
2024/7/31	3.8	4.2	3.4	4.6	4.4	5.1	4.7	5.2	4.7	5.3	20.8	21.3

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師

溫度均在正常範圍內

藥學部 邱保祥 (113) (80)
師(三)級藥師

藥學部 邱子芹 (113) (92)
師(三)級藥師

總藥師:

藥學部 廖志峰 (6802) (1420)
師(二)級藥師

業管主管:

藥學部 何沁沁 (2806) (0930)
科主任