

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 08月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007							
0850-0835	0520-293	1504-0339	1213-0010	1330-0064	1405-0194							
2~8°C	2~8°C	2~8°C	負15~負25°C	≤負70°C	15~25°C							
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/8/1	3.9	4.6	3.7	4.3	3.8	4.3	-21.1	-19.2	-81.8	-80.1	19.7	20.1
2024/8/2	4	4.5	3.7	4.4	3.7	4.3	-21.1	-20.4	-81.9	-79.4	19.7	20.3
2024/8/3	3.9	4.3	3.7	4.4	3.8	4.2	-20.9	-20.7	-81.8	-80.2	19.8	19.9
2024/8/4	3.9	4.3	3.7	4.3	3.8	4.3	-21.1	-20.7	-81.8	-79.7	18.4	19.8
2024/8/5	3.9	4.4	3.7	6.3	3.4	4.3	-21.3	-20.7	-82	-79.2	18.4	13.9
2024/8/6	3.9	4.7	3.7	4.4	3.8	4.3	-21.2	-20.4	-82	-80.2	18.3	19
2024/8/7	4	5.4	3.7	4.4	3.8	4.3	-21.2	-20.6	-81.9	-80.2	18.4	18.9
2024/8/8	4	4.4	3.8	4.4	3.9	4.5	-21.2	-20.4	-82	-79.4	18.4	18.9
2024/8/9	4.1	4.4	3.8	4.5	3.7	4.6	-21.1	-19.9	-81.9	-80.2	18.6	19.3
2024/8/10	4.1	4.3	3.9	4.4	3.8	4.2	-20.9	-20.4	-82	-80.2	18.4	19
2024/8/11	4.2	4.3	3.9	4.4	3.8	4.2	-21.1	-20.7	-82	-78.9	18.9	19.5
2024/8/12	3.7	4.6	3.8	4.4	3.7	4.3	-21.2	-20.4	-82	-80	18.6	19.5
2024/8/13	3.9	5.2	3.9	4.4	3.8	4.3	-21.4	-20.9	-81.9	-79.3	18.4	18.8
2024/8/14	4.1	4.6	3.9	4.4	3.7	4.4	-21.4	-21.1	-81.8	-72.7	18.4	19
2024/8/15	4.2	4.8	3.9	4.4	3.7	4.2	-21.4	-20.8	-81.2	-80.1	18.4	19.5
2024/8/16	4.3	4.9	3.7	4.5	3.7	4.5	-21.4	-20.8	-81.2	-80.1	18.7	19.8
2024/8/17	4.4	6.2	3.7	4.2	3.7	4.2	-21.2	-20.9	-81.4	-80.1	18.6	19.7
2024/8/18	4.4	4.8	3.7	4.2	3.7	4.1	-21.2	-20.7	-81.2	-79.8	19.6	19.7
2024/8/19	4.2	4.8	3.7	4.3	3.7	4.3	-21.1	-20.7	-81.4	-80.2	18.9	19.9
2024/8/20	3.8	4.7	3.7	4.2	3.6	4.2	-21.1	-20.6	-81.4	-79.6	18.6	19.1
2024/8/21	3.8	4.7	3.7	4.3	3.3	4.3	-21.2	-20.7	-81.4	-80.2	18.4	19.4
2024/8/22	4.2	5.1	3.4	4.3	3.5	4.6	-21.5	-20.8	-81.3	-80.2	19.1	19.6
2024/8/23	4.6	5.3	3.4	4.2	3.6	4.2	-21.6	-20.8	-81.4	-80.2	19.2	19.6
2024/8/24	4.7	5.1	3.5	4.2	3.6	4.2	-21.1	-20.8	-81.4	-80.2	19.1	19.4
2024/8/25	4.7	5	3.6	4.2	3.6	4.1	-21.1	-20.8	-81.4	-80.2	19	19.1
2024/8/26	4.6	5.3	3.6	4.2	3.6	4.3	-21.2	-20.9	-81.4	-80.2	19	19.5
2024/8/27	4.4	4.9	3.4	4.2	3.6	4.3	-21.2	-20.4	-81.4	-79	19.1	19.6
2024/8/28	4.3	4.8	3.4	4.2	3.6	4.3	-21.1	-20.4	-81.3	-80.2	19.2	19.7
2024/8/29	3.8	5.1	3.3	4.1	3.8	4.3	-21.1	-20.3	-81.5	-80.3	18.8	19.7
2024/8/30	4	4.8	3.3	4.2	3.8	4.4	-21.4	-20.8	-81.5	-80.3	18.6	19.1
2024/8/31	4	4.7	3.4	4.2	3.9	4.3	-21.4	-21.1	-81.5	-80.2	18.4	13.7

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 113
師(三)級藥師 4902
藥學部 邱子堯 113
師(三)級藥師 01903

總藥師:

藥學部 廖志峰 113
師(二)級藥師 904
0870

業管主管:

藥學部 何沁沁 113
科主任 0706
1350