

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113 年 08 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/8/1	3.7	4.2	3.5	5.1	4.3	4.9	4.7	5.2	4.6	5.2	21.2	21.5
2024/8/2	3.7	4.2	3.5	4.7	4.3	5.1	4.7	5.2	4.5	5.3	21.1	21.5
2024/8/3	3.9	4.1	3.6	4.7	4.3	4.9	4.7	5.2	4.5	5.1	21.3	21.4
2024/8/4	3.9	4.2	3.7	4.7	4.3	4.9	4.7	5.2	4.4	5.1	20.1	21.3
2024/8/5	3.7	4.4	3.6	4.8	4.3	5	4.7	5.2	4.4	5.1	19.9	20.4
2024/8/6	3.7	4.3	3.6	5.1	4.3	4.9	4.7	5.3	4.5	5.7	19.9	20.3
2024/8/7	3.7	4	3.6	4.9	4.4	4.9	4.7	5.4	4.6	5.3	19.9	20.3
2024/8/8	3.7	4.1	3.7	4.8	4	5	4.8	5.5	4.2	5.4	20.1	20.5
2024/8/9	3.7	4.1	3.7	4.8	4	4.6	4.8	5.4	4	5	20.2	20.6
2024/8/10	3.7	4.1	3.7	4.8	4	4.6	4.9	5.4	4	4.7	20.1	20.3
2024/8/11	3.7	4.1	3.7	4.8	4	4.6	4.9	5.3	4	4.7	20.3	20.7
2024/8/12	3.7	4.4	3.8	4.8	3.9	4.7	4.9	5.7	3.7	4.7	20.2	20.7
2024/8/13	3.8	4.5	3.7	4.8	4.1	4.6	5.3	5.8	3.7	4.4	19.9	20.2
2024/8/14	4.1	4.5	3.7	4.8	4.1	4.6	5.3	5.7	3.6	5.4	19.8	20.2
2024/8/15	3.9	4.9	3.6	4.8	3.9	5.2	5.1	5.7	4	5.2	19.9	20.4
2024/8/16	3.9	4.4	3.6	4.8	4.1	4.7	5.1	5.5	4.5	5.6	20.4	21
2024/8/17	4	4.3	3.6	4.8	4.1	4.6	5.1	5.4	5.1	5.6	20.3	20.9
2024/8/18	4	4.4	3.6	4.8	4.1	4.6	5.1	5.4	5.1	5.6	20.8	20.9
2024/8/19	3.8	4.3	3.6	4.8	4.1	4.8	4.9	5.7	5.1	5.7	20.6	20.9
2024/8/20	3.8	4.5	3.6	4.8	4	4.6	4.9	5.3	5.1	5.7	20.1	20.6
2024/8/21	3.9	4.3	3.5	4.8	3.9	4.5	4.9	5.3	4.7	5.8	20.1	20.9
2024/8/22	3.9	4.4	3.6	4.8	3.9	4.5	4.9	5.4	4.7	5.4	20.6	21.1
2024/8/23	3.6	4.3	3.5	4.8	3.9	4.5	4.4	5.4	4.9	5.4	20.7	21.2
2024/8/24	3.7	4.3	3.5	4.8	3.9	4.5	4.6	4.9	4.9	5.4	20.6	20.9
2024/8/25	3.7	4.3	3.5	4.7	3.9	4.5	4.6	5	4.9	5.4	20.5	20.6
2024/8/26	3.7	4.4	3.4	4.8	3.9	4.5	4.6	5.1	4.9	5.8	20.4	20.9
2024/8/27	3.8	4.3	3.4	5.1	3.9	4.6	4.7	5.2	5.2	6.1	20.6	20.8
2024/8/28	3.8	4.6	3.6	5.1	3.9	4.5	4.7	5.6	5.2	5.9	20.6	20.9
2024/8/29	3.7	4.4	3.6	4.8	3.9	4.9	4.9	5.6	5.5	6.1	20.6	21.1
2024/8/30	3.7	4.6	3.6	4.8	4.1	4.8	4.9	5.3	5.7	6.1	20.4	20.7
2024/8/31	3.9	4.4	3.7	4.8	4	4.7	4.9	5.3	5.7	5.9	20.2	20.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 邱保祥 113
師(三)級藥師 1902
藥學部 邱子芹 113
師(三)級藥師 1903

總藥師:

藥學部 廖志峰 0704
師(二)級藥師 0810

業管主管:

藥學部 何沁沁 0904
科主任 1380