

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 09月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362	0520-292	1504-0305	1405-0193	1215-0006	1213-0007							
0850-0835	0520-293	1504-0339	1213-0010	1330-0064	1405-0194							
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/9/1	4	4.7	3.5	4.2	3.9	4.3	-21.4	-21.1	-81.5	-80.3	18.4	18.9
2024/9/2	4.2	5	3.4	4.2	4.1	4.7	-20.9	-18.7	-81.2	-80.1	19	19.4
2024/9/3	4.1	4.8	3.4	4.2	4	4.6	-20.9	-20.3	-81.2	-80.1	19	19.6
2024/9/4	4.1	4.7	3.5	4.2	3.9	4.5	-20.7	-20	-81.2	-79.2	19.1	19.6
2024/9/5	4.1	4.6	3.5	4.3	3.9	4.4	-20.6	-20.2	-81.2	-80.1	18.6	19.2
2024/9/6	4.1	4.5	3.5	4.2	3.9	4.5	-20.8	-20.3	-81.4	-80.2	18.4	18.9
2024/9/7	4.1	4.4	3.5	4.2	4.2	4.5	-20.9	-20.6	-81.4	-80.2	18.3	13.5
2024/9/8	4.1	4.4	3.5	4.2	4.2	4.5	-20.8	-20.6	-81.4	-80.1	18.1	13.3
2024/9/9	4	5.1	3.6	4.3	4.1	4.7	-20.9	-20.4	-81.4	-80	18.1	13.7
2024/9/10	4	5.2	3.5	4.3	3.7	4.6	-20.9	-20.2	-81.2	-80.1	18.2	13.6
2024/9/11	4.4	5.2	3.4	4.3	3.7	4.2	-20.9	-19.9	-81.2	-79.9	18.2	13.8
2024/9/12	4.6	5.7	3.4	4.3	3.7	4.2	-21.4	-20.4	-81.2	-80.1	18.3	13.8
2024/9/13	4.8	5.8	3.5	4.2	3.7	4.3	-21.4	-20.9	-81.2	-73.9	18.3	19.3
2024/9/14	5.1	5.5	3.5	4.1	3.8	4.3	-21.4	-20.9	-80.9	-79.4	18.4	13.6
2024/9/15	5.1	5.5	3.5	4.2	3.8	4.2	-21.4	-21.1	-80.9	-79.3	18.2	18.4
2024/9/16	4.8	5.8	3.6	4.3	3.8	4.4	-21.5	-21.1	-80.9	-79.4	18.2	19.3
2024/9/17	5.3	5.7	3.7	4.2	4	4.3	-21.7	-21.3	-80.9	-79.4	18.8	13.9
2024/9/18	4.6	6	3.7	4.3	3.9	4.7	-21.5	-20.8	-80.6	-78.4	18.8	19.9
2024/9/19	3.7	5.6	3.7	4.3	3.8	4.3	-21	-20.3	-80.7	-79.4	19.5	20.2
2024/9/20	3.3	4.6	3.7	4.3	3.8	4.3	-20.9	-20.6	-80.8	-79.4	19.7	20.1
2024/9/21	3.3	4.3	3.7	4.3	3.9	4.3	-20.9	-20.7	-80.8	-79.2	19.7	19.7
2024/9/22	3.3	4.3	3.7	4.3	3.9	4.3	-20.9	-20.7	-80.9	-79.4	19.5	19.7
2024/9/23	3.3	4.6	3.7	4.3	3.8	5.1	-21.2	-20.7	-80.9	-78.4	19.1	19.9
2024/9/24	3.4	4.2	3.7	4.5	3.9	4.7	-21.2	-20.7	-80.9	-79.4	19.1	20.2
2024/9/25	3.3	4.9	3.7	4.3	3.8	4.3	-21.2	-20.3	-80.8	-79.4	18.8	20.4
2024/9/26	3.2	4.2	3.7	4.4	3.8	4.3	-21.2	-20.7	-80.8	-77.9	18.4	13.9
2024/9/27	3.1	4.7	3.8	4.4	3.8	4.3	-21.1	-20.5	-80.8	-79.4	18.2	19.3
2024/9/28	3.1	4	3.9	4.4	3.8	4.2	-21.1	-20.8	-80.8	-79.5	18.8	13.8
2024/9/29	3.1	4	3.9	4.4	3.8	4.2	-21	-20.8	-80.8	-77.5	18.8	13.8
2024/9/30	3.1	4.1	3.9	4.7	3.8	4.3	-21	-20.7	-80.9	-79.4	18.7	19.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 113
師(三)級藥師 1004 藥學部 邱子萍 113
師(三)級藥師 1074

總藥師: 藥學部 廖志峰 113
師(二)級藥師 0900

業管主管: 藥學部 何沁沁 104
科主任 0960