

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 11月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫I	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/11/1	3.4	4.2	3.9	4.4	3.9	4.6	-20.8	-16.7	-79.7	-77.7	19.4	19.8
2024/11/2	3.4	4.2	3.9	4.4	3.8	4.4	-20.8	-20.5	-79.6	-77.9	19.1	19.4
2024/11/3	3.4	4.2	3.9	4.3	3.8	4.4	-20.8	-20.5	-79.6	-77.9	18.9	19.4
2024/11/4	3.3	4.3	3.9	4.4	3.8	4.4	-20.8	-20.3	-79.6	-78.3	18.8	19.4
2024/11/5	2.8	6.6	3.2	6.5	3.6	6.8	-20.6	-19.9	-79.4	-78.3	19	19.6
2024/11/6	3.6	4.3	3.2	4	3.4	4.1	-20.4	-19.9	-79.4	-78.3	19.1	19.6
2024/11/7	3.7	4.2	3.2	3.9	3.5	4.3	-20.7	-20.2	-79.4	-78.3	19.1	19.6
2024/11/8	3.7	5.1	3.1	4.9	3.6	4.3	-20.7	-19.9	-79.6	-77.9	19.2	19.7
2024/11/9	3.5	4.1	3.1	3.8	3.6	4.2	-20.3	-20.1	-79.6	-77.7	19.2	19.7
2024/11/10	3.6	4.1	3.1	3.7	3.7	4.2	-20.3	-20	-79.6	-77.8	19.2	20.1
2024/11/11	3.7	4.7	3.1	4.4	3.7	4.3	-20.7	-20.1	-79.6	-77.8	18.8	19.6
2024/11/12	3.8	4.3	3.7	4.3	3.5	4.3	-20.7	-19.8	-79.5	-77.4	18.8	20.3
2024/11/13	3.9	4.3	3.7	4.2	3.5	4.3	-20.4	-19.7	-79.4	-78.3	19.4	19.8
2024/11/14	3.9	4.5	3.7	4.1	3.6	4.3	-20.4	-19.9	-79.4	-77.9	19.3	19.9
2024/11/15	3.9	4.3	3.7	4.2	3.7	4.3	-20.2	-19.5	-79.4	-78.3	19.2	19.7
2024/11/16	3.9	4.2	3.8	4.2	3.8	4.3	-20.2	-19.9	-79.6	-77.9	19.1	19.3
2024/11/17	3.9	4.3	3.8	4.1	3.9	4.3	-20.2	-19.9	-79.6	-77.7	18.9	19.1
2024/11/18	3.9	4.5	3.7	4.4	3.7	4.3	-20.4	-19.9	-79.6	-78.3	18.8	19.3
2024/11/19	4	4.5	3.4	4.1	3.7	4.4	-20.4	-19.9	-79.6	-77.2	18.8	19.3
2024/11/20	4.2	4.5	3.4	3.9	3.1	4.3	-20.2	-19.6	-79.4	-78	18.9	19.4
2024/11/21	4.2	5.7	3.4	4.1	3.5	4.2	-20.2	-19.6	-79.6	-78.3	19.2	20.1
2024/11/22	4.1	4.7	3.6	4.2	3.5	4.2	-19.9	-19.3	-79.7	-76.9	19.7	20.1
2024/11/23	4	4.3	3.7	4.1	3.6	4.1	-20.2	-19.9	-79.7	-78.3	19.6	19.7
2024/11/24	4.1	4.3	3.7	4.1	3.6	4.2	-20.2	-20	-79.7	-77.2	19.5	19.6
2024/11/25	3.8	4.6	3.7	4.2	3.4	4.2	-20.7	-19.7	-79.7	-77	19.2	19.9
2024/11/26	3.7	4.4	3.6	4.2	3.1	4.1	-20.7	-20.2	-79.7	-77.2	18.8	19.3
2024/11/27	3.7	4.2	3.5	4.2	3.2	3.9	-20.6	-20.1	-80.8	-76.8	18.6	19
2024/11/28	3.7	4.4	3.5	4	3.2	3.9	-20.6	-19.8	-80.8	-79.4	18.6	19.1
2024/11/29	3.8	4.3	3.4	4.4	3.2	3.9	-20.7	-20	-80.8	-79.2	18.3	18.8
2024/11/30	3.7	4.1	3.4	3.9	3.4	3.9	-20.7	-20.4	-80.6	-79.2	18.4	18.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 113 藥學部 邱保祥 1202 藥學部 邱保祥 113 藥學部 邱保祥 1203

總藥師: 藥學部 廖志峰 113 藥學部 廖志峰 1204 藥學部 廖志峰 0830

業管主管: 藥學部 何沁 1204 科主任 何沁 1800