

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113年 12月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/12/1	3.5	4	3.4	3.9	3.6	4.1	-20.7	-20.4	-78.9	-76.9	19.3	19.4
2024/12/2	3.3	4.1	3.4	4	3.6	4.2	-20.9	-18.5	-79.2	-77.5	19.4	19.6
2024/12/3	3.3	4.1	3.4	3.9	3.4	4.2	-20.9	-20.2	-79.1	-76.6	19.4	19.7
2024/12/4	3.5	4.2	3.4	3.9	3.4	4	-20.9	-20.7	-79.1	-77.3	19.5	19.8
2024/12/5	3.6	4.6	3.4	4	3.3	4.1	-20.9	-20.4	-79.2	-76.5	19.5	19.9
2024/12/6	3.8	4.8	3.5	4	3.6	4.2	-20.9	-20	-79.1	-76.7	19.6	19.8
2024/12/7	3.8	4.4	3.4	3.9	3.6	4.2	-20.9	-20.7	-79.1	-76.8	19.5	19.7
2024/12/8	3.8	4.4	3.5	3.9	3.6	4.2	-20.9	-20.7	-79.1	-76.7	19.2	19.5
2024/12/9	3.8	4.7	3.4	4	3.6	4.2	-20.9	-20.4	-79.1	-76.7	19.2	19.5
2024/12/10	4.1	4.7	3.4	4.1	3.5	4.1	-20.9	-20.3	-79.1	-76.8	19.4	19.9
2024/12/11	4.2	5.9	3.4	4.2	3.2	4.6	-20.9	-20	-79.2	-77.9	19.7	19.9
2024/12/12	4.1	4.9	3.9	4.4	3.2	4.1	-20.9	-20.6	-79.1	-77.9	19.6	20
2024/12/13	4	5.2	3.9	4.5	3.2	3.8	-20.9	-20.6	-79.1	-77.4	19.9	20.1
2024/12/14	4.2	4.5	3.8	4.2	3.2	3.9	-20.9	-20.8	-79.1	-76.9	19.4	19.9
2024/12/15	4.2	4.5	3.8	4.2	3.3	3.9	-20.9	-20.8	-79.1	-76.4	19.3	19.4
2024/12/16	3.8	4.7	3.9	4.4	3.1	3.9	-20.9	-20.7	-79.1	-76.2	19.2	19.4
2024/12/17	3.7	4.5	3.8	4.4	3.1	4.4	-20.9	-20.4	-79.1	-76.4	19.4	19.8
2024/12/18	3.9	5	3.8	4.6	3.3	3.8	-20.9	-20.5	-79.1	-76.9	19.6	19.7
2024/12/19	3.3	5.2	3.9	4.7	3.3	3.9	-21.5	-20.1	-79.3	-75	19.4	19.7
2024/12/20	3.5	4.4	4	5.5	2.9	4.1	-21.2	-20.7	-79.2	-77.9	19.6	22.4
2024/12/21	3.8	4.4	4	4.3	2.9	3.7	-21	-20.8	-79.3	-77.9	19.7	19.9
2024/12/22	3.8	4.4	4	4.3	2.9	3.7	-21	-20.9	-79.3	-77.9	19.3	19.7
2024/12/23	3.7	4.6	3.2	4.5	2.9	4	-21	-20.4	-79.3	-77.7	19.2	19.5
2024/12/24	3.3	4.5	3.7	4.3	3.2	4.7	-20.8	-19	-79.3	-77.6	19.4	19.7
2024/12/25	3.9	4.8	3.9	4.4	3.9	4.4	-20.8	-20.4	-79.3	-76.9	19.7	20.2
2024/12/26	4.1	4.8	3.9	4.3	3.9	4.4	-20.8	-20	-78.8	-77.2	20.1	20.4
2024/12/27	3.9	4.7	3.9	4.3	3.9	4.4	-20.3	-19.6	-78.6	-77.6	19.9	20.1
2024/12/28	4.1	4.7	3.9	4.3	3.9	4.4	-20.3	-20.1	-78.7	-77	19.2	19.9
2024/12/29	4.1	4.7	3.9	4.3	3.9	4.4	-20.3	-20.2	-78.7	-76.8	19	19.2
2024/12/30	3.8	4.8	3.3	4.4	3.8	4.4	-20.3	-20.1	-78.6	-76.8	19.2	19.7
2024/12/31	3.7	4.4	3.9	4.3	3.8	4.5	-20.4	-19.9	-78.6	-76.7	19.2	19.6

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 藥學部 邱子芹

總藥師: 藥學部 廖志峰

業管主管: 藥學部 何沁沁

溫度均在正常範圍內