

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 113 年 12 月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2024/12/1	4.7	4.9	4.2	4.8	4	4.4	4.8	5.1	4.4	5.1	19.9	20.1
2024/12/2	4.7	5.1	4.3	4.9	4	4.4	4.9	5.4	4.4	5.6	19.9	20.2
2024/12/3	4.9	5.2	4.3	4.8	3.9	4.7	4.9	5.3	4.4	4.9	20.1	20.3
2024/12/4	4.9	5.2	4.3	4.8	3.9	4.6	4.9	5.3	4.4	4.9	20.1	20.3
2024/12/5	4.9	5.2	4.3	4.7	3.9	4.5	4.9	5.2	3.7	5.3	20.1	20.4
2024/12/6	4.9	5.2	4.3	4.7	3.9	4.6	4.9	5.3	3.4	4.6	20.1	20.3
2024/12/7	4.9	5.1	4.3	4.7	4.3	4.5	4.9	5.2	3.5	4.4	20.1	20.3
2024/12/8	4.9	5.1	4.3	4.7	4.3	4.6	4.9	5.2	3.5	4.4	19.6	20.1
2024/12/9	4.9	5.2	4.3	4.7	4.2	4.6	4.7	5.3	3.3	4.4	19.6	19.9
2024/12/10	4.4	5.1	4.2	4.7	4.2	5.2	4.7	4.9	3.3	4.6	19.9	20.3
2024/12/11	4.4	4.7	4.2	4.8	3.7	4.5	4.7	5	3.4	4.6	20.1	20.4
2024/12/12	4.4	4.8	4.2	5.6	3.7	5.9	4.4	5.3	3.2	4.7	20.1	20.5
2024/12/13	4.4	5.1	4.2	4.7	3.9	4.4	4.4	4.8	3.6	4.7	20.5	20.8
2024/12/14	4.7	5	4.2	4.7	3.9	4.4	4.4	4.7	3.6	4.6	19.9	20.4
2024/12/15	4.7	4.9	4.2	4.7	3.9	4.4	4.4	4.7	3.6	4.6	19.7	19.9
2024/12/16	4.7	5	4.2	4.7	3.9	5.2	4.3	4.7	3.6	4.6	19.6	19.9
2024/12/17	4.4	5.1	4.1	4.7	4.3	4.8	4.2	4.7	3.6	4.7	19.9	20.3
2024/12/18	4.4	4.7	4.1	4.7	4.3	5.1	4.2	4.7	3.9	4.8	19.9	20.2
2024/12/19	4.3	4.7	4.2	5.1	4.4	5.6	4.2	4.8	3.9	4.9	19.7	20.1
2024/12/20	4.4	4.9	4	4.8	4.4	5.3	4.3	4.8	4.1	4.9	20.1	20.5
2024/12/21	4.4	4.8	4	4.8	4.3	4.8	4.3	4.7	4.1	4.9	19.9	20.3
2024/12/22	4.4	4.8	4	4.8	4.3	4.7	4.3	4.7	4.1	4.9	19.5	19.9
2024/12/23	3.9	5	4.1	4.9	4.3	4.9	4.3	4.7	4	4.9	19.5	19.9
2024/12/24	3.9	4.4	4	4.8	4.3	5.2	4.3	5	4	4.8	19.8	20.2
2024/12/25	4.1	4.3	4	4.9	4.3	5.2	4.3	4.6	3.9	4.9	20.1	20.5
2024/12/26	4	4.3	4.1	5.4	4.3	5.4	4.3	4.9	3.8	4.8	20.3	20.7
2024/12/27	4.2	4.4	4.2	4.8	4.4	4.9	4.6	4.9	3.9	4.8	20.1	21.1
2024/12/28	4.3	4.4	4.2	4.7	4.4	4.8	4.6	4.9	3.9	4.9	19.5	20.2
2024/12/29	4.3	4.4	4.2	4.7	4.4	4.8	4.6	4.9	3.9	4.9	19.3	19.5
2024/12/30	4.3	4.6	4.2	4.7	4.4	4.9	4.5	4.9	3.7	4.9	19.4	19.9
2024/12/31	4.4	4.7	4.2	5.1	4.5	7.1	4.6	4.9	3.6	4.8	19.5	19.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師: 藥學部 邱保祥 114
師(三)級藥師 邱子芹 114

總藥師: 藥學部 廖志峰
師(三)級藥師

業管主管: 藥學部 何沁沁 0103
科主任 700