

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114年 01月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8°C		2~8°C		2~8°C		負15~負25°C		≤負70°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/1/1	3.7	4.3	3.9	4.4	3.7	4.2	-20.4	-19.9	-78.6	-76.9	19.2	20.7
2025/1/2	3.7	5.4	3.9	4.5	3.3	3.9	-20.7	-17.1	-80	-78	19.2	19.5
2025/1/3	3.8	4.6	4	4.4	3.3	3.8	-20.7	-20.1	-80	-78.9	18.8	19.2
2025/1/4	3.8	4.3	4	4.3	3.3	3.8	-20.7	-20.4	-80	-78.8	18.7	18.8
2025/1/5	3.8	4.3	4	4.3	3.3	3.8	-20.7	-20.4	-80	-78.4	18.7	18.7
2025/1/6	3.3	4.5	2.9	4.4	3.3	3.9	-20.8	-20.4	-79.9	-77.9	17.9	19.1
2025/1/7	3.4	4.6	3.1	4.4	3.4	4	-22.9	-20.5	-80	-77.6	17.7	18.1
2025/1/8	3.4	4.3	3.9	4.5	3.4	4.1	-22.5	-21.7	-79.9	-77.5	17.7	18.2
2025/1/9	3.5	4.6	3.9	4.3	3.6	4.2	-21.7	-21.2	-80	-77.9	17.6	18.7
2025/1/10	3.4	4.6	3.9	4.3	3.4	4.1	-21.5	-20.8	-80	-78	17.8	18.3
2025/1/11	3.6	4.3	3.9	4.2	3.6	4.1	-21.4	-21.1	-79.9	-77.8	17.9	18.4
2025/1/12	3.6	4.2	3.9	4.2	3.6	4.1	-21.4	-21.1	-79.9	-78.9	18.2	18.4
2025/1/13	3.3	4.3	3.9	4.3	3.5	4.3	-21.5	-20.9	-79.9	-78.2	18.3	19.1
2025/1/14	3.1	4.2	3.9	4.3	3.5	4.1	-21.7	-21	-79.9	-78.9	18.6	19.2
2025/1/15	3.2	4.3	3.9	4.4	3.4	4.3	-21.7	-20.1	-79.9	-77.7	18.8	19.4
2025/1/16	2.7	4.3	3.9	4.3	3.4	4.1	-21.4	-20.6	-80.1	-77.9	18.8	19.2
2025/1/17	3.6	4.3	3.9	4.3	3.7	4.2	-21.4	-20.9	-79.9	-78.8	18.7	19.1
2025/1/18	3.7	4.3	4	4.3	3.7	4.2	-21.4	-21.2	-79.9	-78.5	18.7	18.8
2025/1/19	3.7	4.3	4	4.3	3.6	4.1	-21.4	-21.3	-79.9	-78.8	18.8	18.9
2025/1/20	3.7	4.4	4	4.4	3.6	4.2	-21.5	-20.7	-79.9	-78.8	18.7	19.2
2025/1/21	3.7	4.4	4	4.4	3.6	4.8	-21.4	-20.9	-79.9	-77.6	18.8	19.4
2025/1/22	3.8	4.6	4.1	4.4	3.5	4.6	-21.2	-20.7	-79.9	-77.4	19.2	19.7
2025/1/23	3.5	4.6	4	4.4	3.5	4.2	-20.9	-20.4	-79.9	-78.6	19.4	20
2025/1/24	3.3	4.3	4	4.4	3.6	4.1	-20.8	-20	-79.9	-78.4	18.6	19.9
2025/1/25	3.3	4	4.1	4.2	3.7	4.2	-21.2	-19.5	-79.9	-78.8	18.1	18.6
2025/1/26	3.3	4	4.1	4.2	3.7	4.2	-21	-20.8	-79.9	-77.1	17.9	18.1
2025/1/27	3.3	4	3.9	4.3	3.7	4.2	-21.2	-20.9	-79.9	-78.7	17.2	17.9
2025/1/28	3.3	4	3.9	4.3	3.7	4.2	-21.2	-20.8	-79.9	-78.8	17.1	17.8
2025/1/29	3.3	4	4	4.3	3.7	4.2	-21.1	-20.9	-79.9	-77.9	17.7	18.1
2025/1/30	3.3	4	4	4.3	3.7	4.2	-21	-20.9	-79.9	-78.2	18.1	18.4
2025/1/31	3.3	4	4	4.3	3.7	4.1	-21	-20.8	-79.9	-78.4	18.4	18.8

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

承辦藥師:

溫度均在正常範圍內

藥學部 葉馨 114 0203
 契約醫務管理組員
 藥學部 邱保祥 114 0203
 師(三)級藥師

總藥師:

藥學部 廖志峰 114 1803
 師(二)級藥師

業管主管:

藥學部 何沁沁 0205 1100
 科主任