

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114年 02月

序號	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
0520-362												
0520-292												
1504-0305												
1405-0193												
1215-0006												
0850-0835												
0520-293												
1504-0339												
1213-0010												
1330-0064												
1405-0194												
2~8°C												
2~8°C												
2~8°C												
負15~負25°C												
≤負70°C												
15~25°C												
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/2/1	3.3	4	4	4.3	3.7	4.1	-20.9	-20.8	-79.9	-78.8	18.8	19
2025/2/2	3.3	4	4	4.3	3.6	4.1	-20.9	-20.8	-79.9	-78.2	19	19.1
2025/2/3	2.9	4.2	4	4.4	3.6	4.2	-22.5	-20.6	-79.9	-78.6	19	19.3
2025/2/4	3.3	4.9	4	4.3	3.6	4.2	-22.7	-22.2	-79.9	-77.9	18.7	19.3
2025/2/5	3.9	5.4	4	4.4	3.3	4.4	-22.7	-21.9	-79.9	-77.4	18.6	19.2
2025/2/6	4	4.9	4	4.6	3.3	4.4	-22.7	-22.2	-80.2	-78.6	18.8	19.6
2025/2/7	4.2	4.7	4.3	4.8	3.9	4.7	-22.7	-22.2	-79.9	-77.4	18.8	19.3
2025/2/8	4.3	4.8	4.4	4.8	3.9	4.7	-22.7	-22.2	-79.9	-77.9	18.6	19.3
2025/2/9	4.3	4.7	4.4	4.7	3.8	4.2	-22.7	-22.3	-79.9	-77.6	18.4	18.6
2025/2/10	4.1	4.7	4.4	4.8	3.8	4.2	-22.9	-22.2	-79.9	-78.2	18.1	19.1
2025/2/11	4.3	4.9	4.4	4.8	3.8	4.3	-22.9	-22.6	-79.9	-78.7	18.8	19.3
2025/2/12	3.6	4.8	4.4	5	3.9	4.2	-22.9	-21.9	-79.9	-78.6	19.2	19.7
2025/2/13	3.7	4.5	4.3	5.4	3.6	4.2	-22.5	-22.2	-79.9	-78.6	19.2	19.6
2025/2/14	3.6	4.3	4.4	5.1	3.5	4	-22.7	-22.1	-79.9	-78.6	19.2	20.1
2025/2/15	3.6	4.2	4.7	4.9	3.6	4.1	-22.7	-22.3	-79.9	-77.7	19.4	19.4
2025/2/16	3.6	4.2	4.7	4.9	3.6	4.1	-22.7	-22.2	-79.9	-78.7	19.3	19.4
2025/2/17	3.4	4.4	4.7	4.9	3.5	4.2	-22.7	-22.2	-79.9	-78.6	19	19.6
2025/2/18	3.7	4.7	3.8	4.9	3.5	4.1	-22.7	-21.5	-79.9	-78.6	19.2	19.6
2025/2/19	4.2	4.7	3.8	4.2	3.6	4.4	-21.9	-21.6	-79.9	-78.5	19.2	19.6
2025/2/20	4.3	4.8	3.8	4.2	3.6	4.1	-21.9	-21.1	-79.9	-78.1	19.3	19.8
2025/2/21	4	4.7	3.9	4.3	3.6	4.2	-21.9	-20.8	-79.9	-78.4	19.4	19.7
2025/2/22	4.2	4.6	4	4.3	3.7	4.2	-20.9	-20.8	-79.9	-77.2	19.2	19.4
2025/2/23	4.2	4.6	4	4.2	3.7	4.2	-20.9	-20.8	-79.9	-78.4	19	19.3
2025/2/24	3.9	4.7	3.9	4.3	3.7	4.2	-20.9	-20.8	-79.9	-77.9	18.9	19.3
2025/2/25	3.9	4.4	3.9	4.7	3.7	4.1	-20.9	-20.7	-79.8	-77.9	18.8	19.3
2025/2/26	3.4	4.5	3.9	4.4	3.7	4.2	-20.9	-20.7	-79.9	-78.5	19.2	19.8
2025/2/27	3.3	4.3	3.9	4.3	3.7	4.4	-20.9	-20.4	-79.9	-78.4	19.4	19.8
2025/2/28	3.4	4.1	3.9	4.2	3.8	4.3	-20.7	-20.5	-79.9	-77.9	19.4	19.5

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師: 藥學部 邱保祥 114  
師(三)級藥師 2303  
藥學部 邱子芹 114  
師(三)級藥師 1035

總藥師: 藥學部 廖志峰 114  
師(二)級藥師 0305  
1540

業管主管: 藥學部 何沁沁 0305  
科主任 1730