

# 臺北榮民總醫院藥學部

## 臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114年 02月

	零號冰箱		四號冰箱		五號冰箱		八號冰箱		九號冰箱		室溫2	
序號	2052-0128		1611-0062		1715-0118		2123-0024		2123-0028		1715-0116	
備用	2052-0129		1611-0052		1715-0119		2123-0025		2123-0029		1715-0117	
範圍	2~8°C		2~8°C		2~8°C		2~8°C		2~8°C		15~25°C	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/2/1	4.6	4.8	4.1	4.7	3.2	4	4.7	5.2	3.2	4.1	19.3	19.7
2025/2/2	4.6	4.7	4.1	4.6	3.2	3.9	4.7	5.2	3.1	4.1	19.7	19.8
2025/2/3	4.1	4.8	4	4.7	2.9	4.3	4.5	5.3	3.2	4.1	19.7	20.1
2025/2/4	3.2	4.5	4	4.7	2.9	4.3	4.6	5.3	3.2	4.1	19.6	19.9
2025/2/5	3.3	4.4	4	4.7	2.6	4.1	4.4	5.3	3.2	4.2	19.4	19.6
2025/2/6	3.9	4.7	4.1	5.2	2.6	4.2	4.5	5.3	2.8	4.2	19.5	19.9
2025/2/7	4.2	5.1	4.7	5.2	3.2	4.3	4.7	5.2	3.1	6.3	19.7	19.9
2025/2/8	4.2	4.4	4.7	5.2	3.4	4.7	4.7	5.1	5.9	6.3	19.3	19.7
2025/2/9	4.2	4.4	4.7	5.1	4	4.7	4.7	4.9	5.9	6.1	19	19.3
2025/2/10	4.2	4.8	4.7	5.2	4	4.8	4.4	5.2	5.9	6.6	18.9	19.3
2025/2/11	3.7	4.4	4.7	5.2	3.9	4.7	4.4	4.7	5.8	6.3	19.2	19.7
2025/2/12	3.7	4.2	4.8	5.3	3.7	4.8	4.4	4.8	5.8	6.1	19.7	20.3
2025/2/13	3.5	4.4	4.8	5.3	3.9	4.8	4.5	5.1	5.7	6.1	20.1	20.2
2025/2/14	3.5	4.2	4.8	5.3	2.7	4.8	4.8	5.1	5.3	6.1	19.8	20.3
2025/2/15	3.9	4.1	4.8	5.2	2.8	4.2	4.8	5	5.3	5.6	20.1	20.2
2025/2/16	3.9	4.1	4.8	5.2	2.8	4.2	4.8	5	5.3	5.6	19.9	20.2
2025/2/17	3.8	4.1	4.8	5.3	2.8	4.7	4.8	5.1	5.2	5.6	19.7	20.2
2025/2/18	3.8	4.7	4.8	5.3	3.7	4.7	4.8	5.1	5.2	5.6	19.7	20.1
2025/2/19	4.1	5.1	4.8	5.3	3.6	4.6	4.8	5.2	5.3	5.6	19.7	20.1
2025/2/20	3.7	4.4	4.7	5.3	3.7	4.5	4.7	5.2	5.3	5.7	19.9	20.4
2025/2/21	3.7	4.1	4.7	5.2	3.7	7.4	4.7	5.2	5.3	5.6	20.2	20.4
2025/2/22	3.8	4.1	4.7	5.2	4.3	4.7	4.8	5.1	5.3	5.5	19.9	20.3
2025/2/23	3.8	4.1	4.7	5.2	4.3	4.6	4.9	5.1	5.3	5.6	19.7	19.9
2025/2/24	3.6	4.2	4.7	5.2	4.2	5.1	4.8	5.2	5.3	5.6	19.6	19.8
2025/2/25	3.4	3.9	4.7	5.3	4.2	4.9	4.8	5.2	5.3	5.9	19.5	19.9
2025/2/26	3.5	4.2	4.7	5.3	4.1	4.8	4.5	5.1	5.3	6.5	19.8	20.2
2025/2/27	3.8	4.3	4.7	5.3	4.2	5.1	4.5	5	5.3	5.8	20.1	20.4
2025/2/28	3.8	4.1	4.7	5.2	4.2	4.8	4.7	4.9	5.3	5.4	19.9	20.2

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師

藥學部 邱保祥 (14)  
 師(三)級藥師 邱子芳 (835)

總藥師:

藥學部 廖志峰  
 師(二)級藥師 廖志峰 (1540)

業管主管:

藥學部 何沁沁 (838)  
 科主任 何沁沁 (1730)