

臺北榮民總醫院藥學部

臨床試驗藥局一般藥品冷藏/冷凍/室溫溫度管制表 114年 03月

	一號冰箱		二號冰箱		三號冰箱		負20度冰箱		負80度冰箱		室溫1	
序號	0520-362		0520-292		1504-0305		1405-0193		1215-0006		1213-0007	
備用	0850-0835		0520-293		1504-0339		1213-0010		1330-0064		1405-0194	
範圍	2~8℃		2~8℃		2~8℃		負15~負25℃		≤負70℃		15~25℃	
日期	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max	Min	Max
2025/3/1	3.4	4.1	3.9	4.2	3.8	4.3	-20.7	-20.4	-79.9	-77.2	19.4	19.4
2025/3/2	3.4	4.1	3.9	4.2	3.8	4.3	-20.7	-20.5	-79.9	-78.4	19.4	19.5
2025/3/3	3.7	4.2	3.7	4.4	3.7	4.4	-18.1	-20.3	-78.3	-79.9	19.7	19.9
2025/3/4	3.3	4.6	3.7	4.4	3.7	4.2	-19.9	-20.9	-78.4	-79.9	19.7	20.2
2025/3/5	3.8	4.4	3.7	4.2	3.7	4.2	-20.3	-20.9	-77.8	-79.9	19.7	20.2
2025/3/6	3.7	5.1	3.7	4.2	3.7	4.1	-20.3	-20.7	-78.2	-79.9	19.6	19.9
2025/3/7	3.7	4.4	3.7	4.4	3.7	4.3	-20.3	-20.8	-78.4	-79.8	19.4	19.8
2025/3/8	3.7	4.3	3.7	4.2	3.7	4.1	-20.6	-20.8	-78.2	-79.8	19.2	19.5
2025/3/9	3.7	4.2	3.8	4.2	3.7	4.1	-20.6	-20.8	-77.7	-79.8	19.2	19.3
2025/3/10	3.6	4.3	3.8	4.2	3.6	4.4	-20.4	-20.9	-77.6	-79.8	19.3	19.8
2025/3/11	3.2	4.4	3.8	4.2	3.8	4.2	-20.7	-21.2	-78.2	-79.8	19.4	19.9
2025/3/12	4	4.5	3.8	4.3	3.8	4.3	-20.7	-21.4	-77.2	-79.7	19.7	20.2
2025/3/13	3.8	4.6	3.7	4.6	3.9	4.3	-20	-22	-78.4	-79.7	19.8	20.8
2025/3/14	3.9	4.4	3.7	4.3	3.9	4.7	-19.6	-21.2	-78.2	-79.7	19.9	20.6
2025/3/15	3.9	4.3	3.7	4.1	3.9	4.3	-20.7	-20.9	-77.9	-79.7	19.7	20
2025/3/16	3.9	4.3	3.7	4	3.9	4.3	-20.8	-20.9	-77.8	-79.7	19.1	19.7
2025/3/17	3.7	4.7	3.7	4.1	3.9	4.7	-20.8	-20.9	-77.9	-79.7	18.9	19.7
2025/3/18	3.7	4.4	3.7	4.3	3.9	4.3	-20.7	-21	-77.8	-79.8	18.9	19.5
2025/3/19	3.7	4.9	3.7	4.2	3.9	4.3	-20.2	-20.9	-76.8	-79.8	18.8	19.4
2025/3/20	4.2	5.3	3.7	4.4	3.9	4.4	-20.6	-20.9	-76.5	-79.4	18.4	19.6
2025/3/21	4.4	5.1	3.9	4.4	3.9	4.4	-20.6	-21.2	-75.6	-77.7	19	19.6
2025/3/22	4.5	4.8	3.9	4.2	3.7	4.3	-20.9	-21.2	-73.7	-77	19	19.3
2025/3/23	4.5	4.8	3.9	4.2	3.7	4.2	-20.9	-21.2	-71.7	-74.6	19.1	19.3
2025/3/24	4.5	5.1	3.8	4.3	3.7	4.2	-20.7	-21.2	-71.2	-73.6	19.2	19.8
2025/3/25	4.1	5.1	3.9	4.3	3.7	4.4	-20.8	-21.2	-72.6	-77.9	19	19.6
2025/3/26	3.7	4.6	3.9	4.4	3.9	4.4	-20.8	-21.4	-78	-81.7	19.2	19.7
2025/3/27	3.7	5.3	3.8	4.4	3.7	6.3	-20.6	-21.4	-78.9	-81.7	19.4	20.1
2025/3/28	3.5	4.4	3.8	4.3	3.7	4.2	-20.9	-21.2	-79.6	-81.7	19.2	19.9
2025/3/29	3.8	4.3	3.8	4.3	3.8	4.2	-20.9	-21.2	-79.8	-81.7	18.7	19.3
2025/3/30	3.8	4.3	3.9	4.2	3.8	4.3	-20.9	-21.2	-79.2	-81.7	18.4	18.7
2025/3/31	3.9	4.4	3.8	4.2	3.9	5.2	-20.6	-21.2	-79.3	-81.7	18.4	19.1

註1: 溫度記錄器型號: ESCORT/ Model No.: EI-IN-D-32-L

註2: 溫度記錄器更換電池或送校正期間取備用紀錄器之資料

溫度均在正常範圍內

承辦藥師:

藥學部 葉馨
契約醫務管理組員
11406061
藥學部 邱保祥
師(二)級藥師 11406061

總藥師:

藥學部 廖志峰
師(二)級藥師 11406061
11406061

業管主管:

藥學部 何沁沁
科主任 11406061
0800